
Midmark M11 - v1.0.5
Total Cycles: 4002

Name: RITTER2
Sterilizer ID: V2313228
10/04/2024 02:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.4 F 0.0
2:00 113.2 F 0.0
4:00 125.3 F 0.2
6:00 170.8 F 2.9
8:00 197.7 F 6.3
10:00 203.1 F 10.2
12:00 220.9 F 7.7
14:00 224.7 F 8.8
16:00 235.0 F 13.5
18:00 237.3 F 10.7
20:00 236.0 F 8.8
22:00 249.1 F 15.2
24:00 260.1 F 21.5
26:00 270.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.4 F 29.8
2:00 271.7 F 29.1
3:00 271.4 F 29.5
4:00 272.2 F 29.7

Min 271.1 F 28.9
Max 272.4 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 219.3 F 0.0
10:00 197.6 F 0.0
15:00 190.6 F 0.0
20:00 186.3 F 0.0
25:00 183.5 F 0.0
30:00 182.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:55
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:05:32
CYCLE COMPLETE

10/04/2024 03:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____