

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2587

Name: RITTER 1  
Sterilizer ID: V2178198  
11/21/2022 02:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.0 F 0.0  
2:00 180.9 F 4.2  
4:00 191.2 F 8.8  
6:00 209.0 F 9.6  
8:00 224.7 F 8.1  
10:00 236.4 F 13.4  
12:00 238.6 F 10.6  
14:00 235.4 F 8.7  
16:00 250.6 F 15.6  
18:00 262.4 F 22.4  
20:00 272.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.5 F 29.1  
2:00 271.8 F 28.5  
3:00 271.6 F 28.7  
4:00 272.4 F 29.1

Min 271.2 F 28.2  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.7 F 0.0  
5:00 220.0 F 0.0  
10:00 233.9 F 0.0  
15:00 240.9 F 0.0  
20:00 245.5 F 0.0  
25:00 227.6 F 0.0  
Dry Cycle Aborted

FILLING: 1:19  
HEATING: 20:17  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 25:54  
TOTAL CYCLE: 00:54:46

11/21/2022 03:26 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_