

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1660

Name: RITTER 1  
Sterilizer ID: V2178198  
02/23/2022 11:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.6 F 0.0  
2:00 196.0 F 3.3  
4:00 199.4 F 7.5  
6:00 218.2 F 13.2  
8:00 230.2 F 7.9  
10:00 242.3 F 13.2  
12:00 240.3 F 10.7  
14:00 237.0 F 8.9  
16:00 251.5 F 15.9  
18:00 263.4 F 22.7  
20:00 271.5 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.3 F 28.7  
2:00 271.3 F 28.2  
3:00 272.6 F 28.8  
4:00 271.5 F 28.1

Min 271.2 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.0 F 0.0  
5:00 232.3 F 0.0  
10:00 224.5 F 0.0  
15:00 229.0 F 0.0  
20:00 229.4 F 0.0  
25:00 227.3 F 0.0  
30:00 230.7 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 20:06  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:58:55  
CYCLE COMPLETE

02/23/2022 12:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_