

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1533

Name: RITTER 1  
Sterilizer ID: V2178198  
01/17/2022 01:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.8 F 0.0  
2:00 157.5 F 2.9  
4:00 176.1 F 6.7  
6:00 197.1 F 11.3  
8:00 215.0 F 5.4  
10:00 228.1 F 10.6  
12:00 230.1 F 7.1  
14:00 243.0 F 13.5  
16:00 242.2 F 12.1  
18:00 254.9 F 18.6  
20:00 266.3 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.0  
1:00 272.6 F 29.8  
2:00 271.6 F 28.9  
3:00 272.5 F 29.8  
4:00 272.0 F 29.2

Min 271.4 F 28.9  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.9 F 0.0  
5:00 231.0 F 0.0  
10:00 228.7 F 0.0  
15:00 226.4 F 0.0  
20:00 225.1 F 0.0  
25:00 227.0 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 21:27  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:54  
CYCLE COMPLETE

01/17/2022 02:26 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_