
Midmark M11 - v1.0.5
Total Cycles: 1677

Name: RITTER2
Sterilizer ID: V2313228
11/03/2022 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.7 F 0.0
2:00 181.4 F 3.2
4:00 187.0 F 7.8
6:00 209.5 F 12.4
8:00 226.1 F 8.5
10:00 237.1 F 10.7
12:00 241.6 F 12.2
14:00 240.8 F 11.1
16:00 254.7 F 18.3
18:00 266.8 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.8 F 29.5
2:00 271.2 F 28.8
3:00 272.7 F 29.6
4:00 271.9 F 29.0

Min 271.0 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.3 F 0.0
5:00 219.3 F 0.0
10:00 203.8 F 0.0
15:00 224.5 F 0.0
20:00 220.7 F 0.0
25:00 221.9 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

11/03/2022 10:12 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____