
Midmark M11 - v1.0.5
Total Cycles: 3599

Name: RITTER2
Sterilizer ID: V2313228
05/08/2024 01:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.5 F 0.0
2:00 103.9 F 0.6
4:00 129.5 F 2.7
6:00 160.3 F 5.9
8:00 185.3 F 9.8
10:00 217.9 F 7.1
12:00 227.8 F 7.4
14:00 235.2 F 11.9
16:00 232.3 F 8.2
18:00 242.7 F 12.5
20:00 242.7 F 12.2
22:00 254.9 F 18.6
24:00 266.0 F 25.3
26:00 273.0 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 272.1 F 29.4
2:00 272.5 F 29.6
3:00 272.3 F 29.9
4:00 272.5 F 30.2

Min 271.7 F 29.4
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.6 F 0.0
5:00 238.1 F 0.0
10:00 201.2 F 0.0
15:00 194.2 F 0.0
20:00 186.2 F 0.0
25:00 214.0 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:20
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:04:37
CYCLE COMPLETE

05/08/2024 03:03 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____