
Midmark M11 - v1.0.5
Total Cycles: 1056

Name: RITTER2
Sterilizer ID: V2313228
05/02/2022 01:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.8 F 0.0
2:00 195.1 F 3.8
4:00 204.8 F 9.2
6:00 223.7 F 7.8
8:00 232.4 F 10.8
10:00 234.0 F 8.9
12:00 238.3 F 9.7
14:00 248.3 F 15.1
16:00 260.8 F 22.1
18:00 271.0 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.5 F 29.8
2:00 271.6 F 29.4
3:00 272.5 F 29.8
4:00 271.5 F 29.2

Min 271.5 F 29.2
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 223.3 F 0.0
10:00 231.2 F 0.0
15:00 225.6 F 0.0
20:00 220.7 F 0.0
25:00 218.6 F 0.0
30:00 217.5 F 0.0
Drying Complete

FILLING: 4:01
HEATING: 18:41
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

05/02/2022 02:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____