
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002299
14:18 19/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:07
STERILIZING 3:08
136.0 C 222kPa 3:08
136.2 C 223kPa 3:38
136.2 C 224kPa 4:08
136.1 C 224kPa 4:38
136.4 C 226kPa 5:08
136.3 C 225kPa 5:38
135.9 C 219kPa 6:08
136.4 C 223kPa 6:38
136.4 C 223kPa 6:38
VENTING 6:38
AIR DRYING 7:30
CYCLE COMPLETE 7:57

Digital Signature #
43771E9E122205AF
