
Midmark M11 - v1.0.0
Total Cycles: 4991

Name: RITTER 1
Sterilizer ID: V2178198
10/11/2024 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.2 F 0.0
2:00 115.5 F 0.1
4:00 129.0 F 0.6
6:00 155.3 F 0.9
8:00 179.0 F 3.6
10:00 200.1 F 6.8
12:00 208.9 F 10.0
14:00 219.8 F 13.5
16:00 223.4 F 7.2
18:00 233.3 F 11.2
20:00 232.3 F 7.2
22:00 243.8 F 12.8
24:00 241.2 F 11.0
26:00 252.4 F 16.5
28:00 262.2 F 22.1
30:00 270.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 271.4 F 28.3
2:00 271.7 F 28.6
3:00 272.3 F 29.1
4:00 272.6 F 29.2

Min 271.3 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.7 F 0.0
5:00 198.3 F 0.0
10:00 198.1 F 0.0
15:00 204.7 F 0.0
20:00 214.1 F 0.0
25:00 226.4 F 0.0
30:00 232.0 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:50
STERILIZING: 4:00
VENTING: 3:60
DRYING: 30:00
TOTAL CYCLE: 01:09:36
CYCLE COMPLETE

10/11/2024 01:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____