
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001923
14:36 07/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:37
STERILIZING 3:35
136.0 C 220kPa 3:35
136.4 C 225kPa 4:05
136.4 C 225kPa 4:35
136.3 C 223kPa 5:05
136.2 C 224kPa 5:35
136.3 C 224kPa 6:05
136.4 C 223kPa 6:35
136.4 C 224kPa 7:05
136.4 C 224kPa 7:05
VENTING 7:05
AIR DRYING 7:56
CYCLE COMPLETE 37:06

Digital Signature #
A11D9317495BC92D
