

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001712  
15:45 07/03/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:19  
PRESSURIZING 2:25  
STERILIZING 3:30  
136.0 C 221kPa 3:30  
136.2 C 222kPa 4:00  
136.4 C 224kPa 4:30  
136.0 C 223kPa 5:00  
136.1 C 222kPa 5:30  
136.3 C 223kPa 6:00  
136.4 C 223kPa 6:30  
136.2 C 223kPa 7:00  
136.2 C 223kPa 7:00  
VENTING 7:00  
AIR DRYING 7:50  
CYCLE COMPLETE 8:02

-----  
Digital Signature #  
C372B504A8BAD3FD  
-----