
Midmark M11 - v1.0.0
Total Cycles: 2891

Name: RITTER 1
Sterilizer ID: V2178198
02/13/2023 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.9 F 0.0
2:00 186.9 F 4.5
4:00 198.8 F 9.2
6:00 218.3 F 7.5
8:00 226.7 F 9.3
10:00 235.7 F 9.4
12:00 242.2 F 12.4
14:00 241.0 F 11.0
16:00 254.3 F 17.8
18:00 265.6 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.2
2:00 271.5 F 28.5
3:00 272.7 F 29.4
4:00 272.0 F 28.8

Min 271.5 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 235.6 F 0.0
10:00 226.7 F 0.0
15:00 230.8 F 0.0
20:00 228.9 F 0.0
25:00 229.4 F 0.0
30:00 230.5 F 0.0

Drying Complete

FILLING: 1:40
HEATING: 19:43
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

02/13/2023 01:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____