
Midmark M11 - v1.0.5
Total Cycles: 3696

Name: RITTER2
Sterilizer ID: V2313228
06/13/2024 02:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.0 F 0.0
2:00 132.1 F 1.7
4:00 153.9 F 4.6
6:00 175.6 F 8.1
8:00 197.1 F 12.2
10:00 220.7 F 5.0
12:00 229.3 F 9.0
14:00 240.2 F 13.4
16:00 236.4 F 10.1
18:00 236.5 F 9.1
20:00 247.2 F 14.3
22:00 257.9 F 20.3
24:00 268.5 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.9 F 29.6
2:00 271.4 F 29.0
3:00 272.9 F 29.7
4:00 271.6 F 29.0

Min 271.3 F 28.9
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 198.1 F 0.0
10:00 192.9 F 0.0
15:00 190.6 F 0.0
20:00 187.9 F 0.0
25:00 185.0 F 0.0
30:00 183.0 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 25:14
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:03:45
CYCLE COMPLETE

06/13/2024 03:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____