
Midmark M11 - v1.0.0
Total Cycles: 911

Name: RITTER 1
Sterilizer ID: V2178198
06/17/2021 01:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.0 F 0.0
2:00 138.8 F 1.2
4:00 156.0 F 3.4
6:00 178.2 F 6.5
8:00 197.6 F 10.6
10:00 224.3 F 5.5
12:00 226.3 F 8.8
14:00 237.8 F 13.5
16:00 238.1 F 10.4
18:00 235.4 F 8.3
20:00 248.6 F 14.4
22:00 260.0 F 20.7
24:00 270.2 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 271.5 F 28.6
2:00 271.7 F 28.5
3:00 272.0 F 28.7
4:00 272.1 F 28.8

Min 271.4 F 28.5
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 244.7 F 0.0
10:00 223.3 F 0.0
15:00 232.0 F 0.0
20:00 223.6 F 0.0
25:00 231.3 F 0.0
30:00 216.2 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 25:28
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:03:57
CYCLE COMPLETE

06/17/2021 02:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____