
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003955
8:30 07/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:58
STERILIZING 3:03
136.4 C 222kPa 3:03
136.3 C 222kPa 3:33
136.1 C 220kPa 4:03
136.4 C 224kPa 4:33
136.1 C 221kPa 5:03
136.0 C 221kPa 5:33
136.2 C 220kPa 6:03
136.1 C 218kPa 6:33
136.1 C 218kPa 6:33
VENTING 6:33
AIR DRYING 7:18
CYCLE COMPLETE 7:45

Digital Signature #
7ACBB2FF5A153CC5
