
Midmark M11 - v1.0.0
Total Cycles: 1315

Name: RITTER 1
Sterilizer ID: V2178198
10/28/2021 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.8 F 0.0
2:00 178.7 F 3.0
4:00 193.4 F 6.8
6:00 212.0 F 12.3
8:00 224.6 F 6.5
10:00 238.4 F 11.5
12:00 234.7 F 8.4
14:00 241.0 F 10.6
16:00 246.6 F 13.3
18:00 259.1 F 20.0
20:00 270.0 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.4 F 28.6
2:00 271.8 F 28.4
3:00 272.3 F 28.5
4:00 271.6 F 28.3

Min 271.3 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 195.1 F 0.0
10:00 197.3 F 0.0
15:00 197.4 F 0.0
20:00 194.6 F 0.0
25:00 192.3 F 0.0
30:00 190.0 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:53
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:50
CYCLE COMPLETE

10/28/2021 01:02 PM
ACCEPTED: X
REJECTED:
Operator: U02