
Midmark M11 - v1.0.5
Total Cycles: 1481

Name: RITTER2
Sterilizer ID: V2313228
09/13/2022 07:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.6 F 0.0
2:00 182.7 F 1.5
4:00 194.5 F 5.0
6:00 204.9 F 10.4
8:00 228.4 F 5.8
10:00 235.6 F 11.8
12:00 236.3 F 9.2
14:00 237.3 F 9.6
16:00 247.0 F 14.1
18:00 259.0 F 20.7
20:00 269.8 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.6
1:00 272.0 F 29.9
2:00 272.2 F 29.6
3:00 271.1 F 29.6
4:00 271.2 F 29.9

Min 271.0 F 29.5
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.3 F 0.0
5:00 233.8 F 0.0
10:00 219.7 F 0.0
15:00 202.0 F 0.0
20:00 189.7 F 0.0
25:00 182.2 F 0.0
30:00 172.4 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 21:35
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:39
CYCLE COMPLETE

09/13/2022 08:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____