

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2709

Name: RITTER 1  
Sterilizer ID: V2178198  
12/21/2022 12:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.4 F 0.0  
2:00 186.1 F 3.6  
4:00 197.4 F 8.0  
6:00 217.7 F 13.2  
8:00 228.8 F 8.2  
10:00 240.5 F 12.3  
12:00 239.5 F 11.1  
14:00 238.2 F 9.4  
16:00 252.2 F 16.3  
18:00 264.0 F 23.3  
20:00 271.4 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.6 F 29.0  
2:00 271.5 F 28.2  
3:00 272.4 F 29.1  
4:00 272.2 F 28.7

Min 271.3 F 28.1  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 215.5 F 0.0  
5:00 227.1 F 0.0  
10:00 234.7 F 0.0  
15:00 232.1 F 0.0  
20:00 227.1 F 0.0  
25:00 231.2 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 20:02  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:46  
CYCLE COMPLETE

12/21/2022 01:19 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_