
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005479
12:35 03/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:40
STERILIZING 3:45
136.4 C 220kPa 3:45
136.3 C 225kPa 4:15
136.4 C 225kPa 4:45
136.0 C 221kPa 5:15
136.2 C 221kPa 5:45
136.2 C 223kPa 6:15
136.3 C 223kPa 6:45
136.5 C 225kPa 7:15
136.5 C 225kPa 7:15
VENTING 7:15
AIR DRYING 8:02
CYCLE COMPLETE 20:22

Digital Signature #
A8820FB45CC75CB6
