
Midmark M11 - v1.0.0
Total Cycles: 2062

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2022 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.6 F 0.0
2:00 146.1 F 1.1
4:00 170.0 F 4.1
6:00 200.0 F 7.9
8:00 220.6 F 13.4
10:00 224.6 F 6.8
12:00 236.6 F 11.6
14:00 234.7 F 8.4
16:00 241.4 F 10.9
18:00 246.0 F 13.1
20:00 258.3 F 19.7
22:00 269.5 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.2
1:00 272.3 F 28.9
2:00 271.5 F 28.1
3:00 271.8 F 28.8
4:00 272.1 F 28.7

Min 271.0 F 28.0
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 215.3 F 0.0
10:00 210.0 F 0.0
15:00 206.8 F 0.0
20:00 238.5 F 0.0
25:00 235.1 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:02
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

06/27/2022 02:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____