
Midmark M11 - v1.0.5
Total Cycles: 565

Name: RITTER2
Sterilizer ID: V2313228
11/18/2021 12:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.7 F 0.0
2:00 174.2 F 3.6
4:00 197.9 F 8.1
6:00 214.3 F 11.9
8:00 226.3 F 9.6
10:00 233.9 F 9.2
12:00 243.1 F 13.3
14:00 243.7 F 12.4
16:00 256.0 F 19.1
18:00 267.2 F 26.0
20:00 273.5 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.8
1:00 272.2 F 29.7
2:00 272.2 F 30.0
3:00 272.2 F 30.3
4:00 271.5 F 29.8

Min 271.5 F 29.7
Max 272.8 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 225.5 F 0.0
10:00 223.5 F 0.0
15:00 217.8 F 0.0
20:00 208.0 F 0.0
25:00 221.6 F 0.0
30:00 218.5 F 0.0
Drying Complete

FILLING: 2:31
HEATING: 20:13
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:00:04
CYCLE COMPLETE

11/18/2021 01:12 PM
ACCEPTED: X
REJECTED:
Operator: U02