

-----  
Midmark M11 - v1.0.5  
Total Cycles: 759

Name: RITTER2  
Sterilizer ID: V2313228  
02/03/2022 11:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.2 F 0.0  
2:00 194.3 F 3.8  
4:00 204.2 F 9.2  
6:00 225.6 F 7.8  
8:00 234.7 F 11.1  
10:00 235.2 F 9.3  
12:00 238.0 F 9.4  
14:00 249.8 F 15.6  
16:00 262.1 F 22.7  
18:00 272.8 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.9  
1:00 272.8 F 29.8  
2:00 272.0 F 29.1  
3:00 271.4 F 29.1  
4:00 272.5 F 29.8  
  
Min 271.2 F 28.8  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.4 F 0.0  
5:00 225.3 F 0.0  
10:00 231.7 F 0.0  
15:00 227.1 F 0.0  
20:00 222.8 F 0.0  
25:00 219.9 F 0.0  
30:00 218.9 F 0.0  
Drying Complete

FILLING: 2:37  
HEATING: 18:20  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:03  
CYCLE COMPLETE

02/03/2022 12:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_