

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2512

Name: RITTER 1  
Sterilizer ID: V2178198  
11/03/2022 02:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.1 F 0.0  
2:00 172.9 F 3.1  
4:00 185.5 F 6.7  
6:00 205.4 F 11.6  
8:00 218.9 F 4.7  
10:00 228.7 F 9.4  
12:00 239.4 F 10.1  
14:00 240.3 F 11.3  
16:00 237.7 F 9.2  
18:00 251.4 F 15.8  
20:00 262.6 F 22.3  
22:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 272.4 F 29.2  
2:00 272.4 F 28.8  
3:00 271.3 F 28.2  
4:00 271.5 F 28.7

Min 271.1 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.8 F 0.0  
5:00 224.0 F 0.0  
10:00 241.0 F 0.0  
15:00 236.5 F 0.0  
20:00 229.5 F 0.0  
25:00 231.2 F 0.0  
30:00 227.7 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 22:27  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:00:48  
CYCLE COMPLETE

11/03/2022 03:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_