

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002555  
7:38 28/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:12  
PRESSURIZING 2:16  
STERILIZING 3:17  
136.4 C 223kPa 3:17  
136.0 C 221kPa 3:47  
136.4 C 223kPa 4:17  
136.2 C 222kPa 4:47  
136.3 C 222kPa 5:17  
136.4 C 222kPa 5:47  
135.7 C 218kPa 6:17  
136.3 C 221kPa 6:47  
136.3 C 221kPa 6:47  
VENTING 6:48  
AIR DRYING 7:36  
CYCLE COMPLETE 7:53

-----  
Digital Signature #  
2BCA89709CFA35D7  
-----