
Midmark M11 - v1.0.0
Total Cycles: 1601

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2022 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.4 F 0.0
2:00 139.8 F 2.4
4:00 161.2 F 6.1
6:00 186.5 F 10.6
8:00 211.0 F 4.2
10:00 223.6 F 9.4
12:00 232.3 F 9.5
14:00 239.6 F 11.5
16:00 238.4 F 9.7
18:00 252.2 F 16.5
20:00 264.0 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 29.0
2:00 271.6 F 28.6
3:00 272.5 F 29.0
4:00 271.5 F 28.4

Min 271.4 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 223.9 F 0.0
10:00 236.6 F 0.0
15:00 226.0 F 0.0
20:00 229.8 F 0.0
25:00 229.4 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:56
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

02/07/2022 02:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____