
Midmark M11 - v1.0.0
Total Cycles: 1162

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2021 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.2 F 0.0
2:00 188.6 F 3.0
4:00 200.3 F 7.2
6:00 224.9 F 12.5
8:00 223.4 F 8.0
10:00 237.8 F 13.4
12:00 241.1 F 11.1
14:00 238.2 F 9.4
16:00 251.9 F 16.1
18:00 263.3 F 22.6
20:00 271.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.5 F 28.8
2:00 272.0 F 28.6
3:00 272.0 F 28.6
4:00 272.0 F 28.7

Min 270.7 F 28.0
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 229.5 F 0.0
10:00 227.8 F 0.0
15:00 230.2 F 0.0
20:00 229.3 F 0.0
25:00 228.8 F 0.0
Dry Cycle Aborted

FILLING: 1:55
HEATING: 20:06
STERILIZING: 4:00
VENTING: 3:11
DRYING: 26:36
TOTAL CYCLE: 00:55:47

09/13/2021 11:04 AM
ACCEPTED: X
REJECTED:
Operator: U01