
Midmark M11 - v1.0.0
Total Cycles: 4907

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2024 11:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.2 F 0.0
2:00 179.0 F 0.2
4:00 195.1 F 1.0
6:00 210.4 F 5.1
8:00 219.1 F 9.4
10:00 231.0 F 9.9
12:00 233.8 F 9.4
14:00 238.0 F 9.7
16:00 244.4 F 12.6
18:00 242.6 F 11.3
20:00 254.9 F 17.6
22:00 265.5 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.0
1:00 271.7 F 28.2
2:00 272.5 F 28.8
3:00 272.4 F 28.6
4:00 271.9 F 28.2

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 205.6 F 0.0
10:00 204.1 F 0.0
15:00 207.7 F 0.0
20:00 216.2 F 0.0
25:00 226.3 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 23:53
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:02:28
CYCLE COMPLETE

09/13/2024 12:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____