

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000397  
13:47 25/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:59  
PRESSURIZING 3:02  
STERILIZING 4:16  
136.5 C 226kPa 4:16  
136.3 C 232kPa 4:46  
136.2 C 232kPa 5:16  
136.3 C 232kPa 5:46  
136.1 C 230kPa 6:16  
136.4 C 233kPa 6:46  
136.2 C 231kPa 7:16  
136.4 C 233kPa 7:46  
136.4 C 233kPa 7:46  
VENTING 7:46  
AIR DRYING 8:41  
CYCLE COMPLETE 8:42

-----  
Digital Signature #  
B95C4B74CCFD70C7  
-----