
Midmark M11 - v1.0.0
Total Cycles: 4721

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2024 07:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 71F

** ITEMS NOT STERILE **

07/12/2024 07:57 AM