

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3558

Name: RITTER 1  
Sterilizer ID: V2178198  
07/20/2023 01:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.0 F 0.0  
2:00 143.6 F 0.6  
4:00 164.5 F 3.3  
6:00 189.2 F 7.1  
8:00 205.8 F 11.8  
10:00 222.5 F 6.7  
12:00 230.9 F 12.0  
14:00 234.8 F 8.9  
16:00 238.9 F 9.8  
18:00 246.0 F 13.3  
20:00 257.8 F 19.7  
22:00 269.1 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.0  
1:00 272.6 F 29.2  
2:00 272.8 F 29.6  
3:00 271.6 F 29.0  
4:00 272.7 F 29.3

Min 271.6 F 28.9  
Max 273.1 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
5:00 221.4 F 0.0  
10:00 228.9 F 0.0  
15:00 237.6 F 0.0  
20:00 224.9 F 0.0  
25:00 223.0 F 0.0  
30:00 219.7 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 23:41  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:01:45  
CYCLE COMPLETE

07/20/2023 02:36 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_