

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 005426  
8:01 25/04/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:02  
PRESSURIZING 2:04  
STERILIZING 3:05  
136.5 C 221kPa 3:05  
136.1 C 221kPa 3:35  
136.2 C 222kPa 4:05  
136.3 C 222kPa 4:35  
136.3 C 221kPa 5:05  
136.3 C 223kPa 5:35  
136.4 C 221kPa 6:05  
136.4 C 223kPa 6:35  
136.4 C 223kPa 6:35  
VENTING 6:35  
AIR DRYING 7:25  
CYCLE COMPLETE 7:30

-----  
Digital Signature #  
08C645916D565443  
-----