

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3882

Name: RITTER 1  
Sterilizer ID: V2178198  
10/10/2023 01:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.6 F 0.0  
2:00 174.4 F 3.6  
4:00 203.9 F 7.9  
6:00 227.1 F 13.4  
8:00 229.1 F 8.5  
10:00 239.7 F 12.1  
12:00 239.9 F 11.1  
14:00 237.6 F 9.4  
16:00 251.4 F 16.1  
18:00 263.4 F 22.9  
20:00 271.6 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.3  
1:00 272.9 F 29.0  
2:00 272.6 F 28.9  
3:00 272.4 F 28.8  
4:00 272.6 F 28.9

Min 270.9 F 28.2  
Max 273.0 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.8 F 0.0  
5:00 182.8 F 0.0  
10:00 203.0 F 0.0  
15:00 204.2 F 0.0  
20:00 202.8 F 0.0  
25:00 204.8 F 0.0  
30:00 209.6 F 0.0  
Drying Complete

FILLING: 2:05  
HEATING: 20:06  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 00:59:36  
CYCLE COMPLETE

10/10/2023 02:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_