
Midmark M11 - v1.0.5
Total Cycles: 634

Name: RITTER2
Sterilizer ID: V2313228
12/15/2021 09:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.6 F 0.0
2:00 147.3 F 1.7
4:00 156.3 F 4.8
6:00 182.2 F 9.5
8:00 214.1 F 6.2
10:00 227.4 F 8.0
12:00 240.9 F 13.1
14:00 237.2 F 10.2
16:00 236.4 F 8.4
18:00 249.8 F 15.4
20:00 261.6 F 22.3
22:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 271.8 F 28.8
2:00 272.5 F 29.0
3:00 272.3 F 29.3
4:00 271.6 F 29.2

Min 271.4 F 28.8
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.6 F 0.0
5:00 229.2 F 0.0
10:00 216.4 F 0.0
15:00 224.6 F 0.0
20:00 226.1 F 0.0
25:00 234.2 F 0.0
30:00 232.2 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 22:19
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:00:41
CYCLE COMPLETE

12/15/2021 10:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____