
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000781
9:04 29/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:16
STERILIZING 3:17
136.0 C 228kPa 3:17
136.4 C 232kPa 3:47
136.4 C 233kPa 4:17
136.1 C 231kPa 4:47
136.3 C 232kPa 5:17
136.4 C 233kPa 5:47
136.4 C 231kPa 6:17
136.4 C 231kPa 6:47
136.4 C 231kPa 6:47
VENTING 6:47
AIR DRYING 7:40
CYCLE COMPLETE 7:41

Digital Signature #
1C2BD75C08FA67C3
