
Midmark M11 - v1.0.0
Total Cycles: 4801

Name: RITTER 1
Sterilizer ID: V2178198
08/12/2024 08:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 69.7 F 0.0
4:00 77.8 F 0.1
6:00 102.8 F 0.2
8:00 138.1 F 0.4
10:00 169.6 F 1.0
12:00 191.3 F 3.6
14:00 219.9 F 7.0
16:00 225.8 F 10.9
18:00 226.1 F 6.0
20:00 236.0 F 9.7
22:00 237.4 F 9.7
24:00 243.0 F 11.8
26:00 239.1 F 9.9
28:00 252.6 F 16.4
30:00 263.7 F 22.8
32:00 272.1 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.0
1:00 272.4 F 28.8
2:00 271.8 F 28.0
3:00 272.6 F 28.8
4:00 271.8 F 28.0

Min 271.5 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 222.2 F 0.0
10:00 220.0 F 0.0
15:00 233.6 F 0.0
20:00 227.6 F 0.0
25:00 176.1 F 0.0
30:00 170.5 F 0.0

Drying Complete

FILLING: 0:41
HEATING: 32:01
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:09:59
CYCLE COMPLETE

08/12/2024 09:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____