
Midmark M11 - v1.0.5
Total Cycles: 294

Name: RITTER2
Sterilizer ID: V2313228
08/20/2021 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.4 F 0.0
2:00 73.7 F 0.0
4:00 78.2 F 0.0
6:00 97.6 F 0.2
8:00 136.3 F 2.4
10:00 166.7 F 6.3
12:00 191.5 F 10.8
14:00 206.4 F 5.1
16:00 220.2 F 9.8
18:00 228.6 F 9.2
20:00 236.0 F 12.4
22:00 237.6 F 11.0
24:00 251.0 F 17.6
26:00 263.1 F 24.3
28:00 271.9 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 30.1
1:00 272.6 F 30.9
2:00 271.6 F 30.0
3:00 272.7 F 30.8
4:00 271.7 F 30.0

Min 271.5 F 30.0
Max 272.8 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.5 F 0.0
5:00 232.1 F 0.0
10:00 216.6 F 0.0
15:00 223.1 F 0.0
20:00 215.4 F 0.0
25:00 220.7 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 28:49
STERILIZING: 4:00
VENTING: 2:54
DRYING: 30:00
TOTAL CYCLE: 01:06:53
CYCLE COMPLETE

08/20/2021 09:57 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____