
Midmark M11 - v1.0.5
Total Cycles: 754

Name: RITTER2
Sterilizer ID: V2313228
02/02/2022 03:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.9 F 0.0
2:00 148.6 F 1.3
4:00 170.3 F 5.2
6:00 196.6 F 10.1
8:00 215.5 F 5.8
10:00 228.8 F 10.5
12:00 231.4 F 7.6
14:00 241.0 F 11.5
16:00 245.7 F 13.6
18:00 258.6 F 20.7
20:00 270.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.6 F 29.8
2:00 272.1 F 29.2
3:00 271.3 F 29.0
4:00 272.2 F 29.8

Min 271.3 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 223.8 F 0.0
10:00 206.2 F 0.0
15:00 207.9 F 0.0
20:00 228.5 F 0.0
25:00 222.8 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

02/02/2022 04:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____