
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000615
15:23 27/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:19
PRESSURIZING 2:26
STERILIZING 3:32
136.0 C 229kPa 3:32
136.2 C 232kPa 4:02
136.4 C 233kPa 4:32
136.2 C 231kPa 5:02
136.2 C 231kPa 5:32
136.1 C 231kPa 6:02
136.3 C 232kPa 6:32
136.4 C 233kPa 7:02
136.4 C 233kPa 7:02
VENTING 7:02
AIR DRYING 7:55
CYCLE COMPLETE 22:55

Digital Signature #
BFFCCD424810D682
