
Midmark M11 - v1.0.0
Total Cycles: 999

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2021 06:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.9 F 0.0
2:00 93.0 F 0.0
4:00 95.6 F 0.0
6:00 98.3 F 0.1
8:00 126.3 F 0.1
10:00 163.9 F 0.2
12:00 191.8 F 1.2
14:00 211.4 F 4.4
16:00 221.2 F 6.9
18:00 228.8 F 10.0
20:00 237.4 F 12.3
22:00 237.5 F 9.7
24:00 239.9 F 10.0
26:00 244.6 F 12.4
28:00 241.0 F 10.6
30:00 252.9 F 16.4
32:00 263.2 F 22.4
34:00 272.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.9
1:00 272.6 F 28.6
2:00 271.8 F 27.9
3:00 271.6 F 28.2
4:00 272.5 F 28.5

Min 271.2 F 27.7
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.7 F 0.0
5:00 231.5 F 0.0
10:00 222.8 F 0.0
15:00 219.1 F 0.0
20:00 216.2 F 0.0
25:00 229.2 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 34:26
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:12:31
CYCLE COMPLETE

07/13/2021 07:54 AM
ACCEPTED: X
REJECTED:
Operator: U01