
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000881
16:00 21/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:18
STERILIZING 4:33
136.0 C 227kPa 4:33
136.3 C 232kPa 5:03
136.4 C 233kPa 5:33
136.4 C 231kPa 6:03
136.2 C 232kPa 6:33
136.3 C 232kPa 7:03
136.3 C 232kPa 7:33
136.2 C 230kPa 8:03
136.2 C 230kPa 8:03
VENTING 8:04
AIR DRYING 8:59
CYCLE COMPLETE 23:59

Digital Signature #
C437BB1E3545A0FA
