
Midmark M11 - v1.0.5
Total Cycles: 34

Name: RITTER2
Sterilizer ID: V2313228
05/19/2021 03:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.8 F 0.0
2:00 189.9 F 4.1
4:00 209.2 F 9.7
6:00 221.7 F 6.5
8:00 236.6 F 11.5
10:00 239.1 F 10.5
12:00 239.6 F 10.3
14:00 254.4 F 17.7
16:00 266.6 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.6
1:00 272.7 F 29.3
2:00 271.6 F 28.6
3:00 272.5 F 29.1
4:00 271.6 F 28.7

Min 271.5 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 232.1 F 0.0
10:00 228.4 F 0.0
15:00 227.4 F 0.0
20:00 223.2 F 0.0
25:00 203.5 F 0.0
30:00 203.8 F 0.0

Drying Complete

FILLING: 3:03
HEATING: 17:19
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:27
CYCLE COMPLETE

05/19/2021 04:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____