
Midmark M11 - v1.0.5
Total Cycles: 3238

Name: RITTER2
Sterilizer ID: V2313228
12/20/2023 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.1 F 0.0
2:00 144.5 F 1.1
4:00 151.8 F 4.3
6:00 178.0 F 8.2
8:00 195.5 F 12.8
10:00 217.4 F 6.9
12:00 230.4 F 11.7
14:00 231.2 F 8.7
16:00 239.1 F 10.7
18:00 245.0 F 13.3
20:00 257.2 F 19.8
22:00 268.1 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 272.1 F 29.5
2:00 272.5 F 29.7
3:00 272.9 F 29.9
4:00 272.5 F 30.1

Min 271.7 F 29.5
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 231.0 F 0.0
10:00 229.6 F 0.0
15:00 229.8 F 0.0
20:00 231.4 F 0.0
25:00 228.7 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:56
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:02:26
CYCLE COMPLETE

12/20/2023 02:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____