
Midmark M11 - v1.0.5
Total Cycles: 2140

Name: RITTER2
Sterilizer ID: V2313228
03/13/2023 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.2 F 0.0
2:00 190.8 F 3.4
4:00 208.9 F 7.8
6:00 226.0 F 13.5
8:00 227.2 F 8.3
10:00 236.5 F 11.1
12:00 241.5 F 12.2
14:00 240.9 F 11.2
16:00 254.6 F 18.3
18:00 266.6 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.8 F 30.2
2:00 272.4 F 29.6
3:00 271.6 F 29.3
4:00 272.6 F 30.2

Min 271.6 F 29.2
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.3 F 0.0
5:00 224.1 F 0.0
10:00 238.7 F 0.0
15:00 228.0 F 0.0
20:00 215.4 F 0.0
25:00 217.1 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:36
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

03/13/2023 11:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____