
Midmark M11 - v1.0.5
Total Cycles: 126

Name: RITTER2
Sterilizer ID: V2313228
06/22/2021 11:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.8 F 0.0
2:00 143.1 F 1.4
4:00 145.1 F 3.6
6:00 166.1 F 7.0
8:00 182.6 F 11.5
10:00 213.8 F 4.9
12:00 222.5 F 8.9
14:00 232.8 F 13.1
16:00 235.6 F 10.1
18:00 239.6 F 9.6
20:00 246.1 F 14.2
22:00 257.0 F 20.0
24:00 267.0 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 271.5 F 29.8
2:00 272.4 F 30.0
3:00 271.4 F 29.4
4:00 271.8 F 30.1

Min 270.9 F 29.2
Max 273.0 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.9 F 0.0
5:00 219.4 F 0.0
10:00 225.3 F 0.0
15:00 222.3 F 0.0
20:00 238.6 F 0.0
25:00 228.2 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 25:30
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:04:29
CYCLE COMPLETE

06/22/2021 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____