
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002427
8:52 13/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:50
STERILIZING 3:55
136.0 C 220kPa 3:55
136.4 C 224kPa 4:25
136.4 C 224kPa 4:55
136.4 C 221kPa 5:25
136.4 C 220kPa 5:55
136.0 C 222kPa 6:25
136.4 C 224kPa 6:55
136.2 C 222kPa 7:25
136.2 C 222kPa 7:25
VENTING 7:25
AIR DRYING 8:17
CYCLE COMPLETE 8:18

Digital Signature #
02AD6E51CA1BC701
