

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2206

Name: RITTER 1  
Sterilizer ID: V2178198  
08/08/2022 10:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
2:00 183.6 F 3.4  
4:00 196.9 F 7.9  
6:00 213.6 F 13.0  
8:00 222.8 F 7.6  
10:00 236.2 F 12.8  
12:00 238.3 F 10.0  
14:00 235.9 F 8.4  
16:00 249.3 F 14.7  
18:00 261.2 F 21.4  
20:00 271.9 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.8 F 29.3  
2:00 272.6 F 28.9  
3:00 271.5 F 28.3  
4:00 271.8 F 28.8  
  
Min 271.3 F 28.3  
Max 273.0 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.3 F 0.0  
5:00 240.4 F 0.0  
10:00 235.1 F 0.0  
15:00 229.4 F 0.0  
20:00 228.1 F 0.0  
25:00 227.2 F 0.0  
30:00 230.7 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 21:23  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:45  
CYCLE COMPLETE

08/08/2022 11:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_