
Midmark M11 - v1.0.5
Total Cycles: 526

Name: RITTER2
Sterilizer ID: V2313228
11/09/2021 09:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.8 F 0.0
2:00 100.0 F 0.0
4:00 115.6 F 0.1
6:00 156.1 F 2.8
8:00 180.3 F 6.8
10:00 198.3 F 11.4
12:00 214.7 F 5.6
14:00 225.8 F 10.4
16:00 232.6 F 8.1
18:00 242.5 F 13.1
20:00 242.3 F 12.0
22:00 254.7 F 18.4
24:00 266.0 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.7 F 29.6
2:00 271.3 F 29.0
3:00 272.5 F 29.7
4:00 271.8 F 29.0

Min 271.3 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.5 F 0.0
5:00 202.5 F 0.0
10:00 214.6 F 0.0
15:00 245.2 F 0.0
20:00 229.1 F 0.0
25:00 229.6 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 25:40
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:04:02
CYCLE COMPLETE

11/09/2021 10:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____