

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1981

Name: RITTER 1  
Sterilizer ID: V2178198  
05/31/2022 12:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.9 F 0.0  
2:00 163.2 F 3.4  
4:00 182.2 F 7.9  
6:00 202.2 F 12.9  
8:00 218.1 F 6.3  
10:00 231.9 F 11.3  
12:00 230.8 F 7.4  
14:00 243.5 F 13.4  
16:00 242.3 F 11.8  
18:00 255.4 F 18.4  
20:00 266.7 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.4 F 29.4  
2:00 272.2 F 28.9  
3:00 271.3 F 28.4  
4:00 272.1 F 29.3  
  
Min 271.2 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.7 F 0.0  
5:00 206.3 F 0.0  
10:00 216.5 F 0.0  
15:00 222.0 F 0.0  
20:00 227.3 F 0.0  
25:00 229.5 F 0.0  
30:00 228.2 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 21:35  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:59  
CYCLE COMPLETE

05/31/2022 02:23 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02