

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 003803  
8:55 09/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:41  
STERILIZING 3:45  
136.4 C 223kPa 3:45  
136.1 C 222kPa 4:15  
136.2 C 224kPa 4:45  
136.4 C 224kPa 5:15  
136.4 C 223kPa 5:45  
136.2 C 224kPa 6:15  
136.4 C 222kPa 6:45  
136.1 C 221kPa 7:15  
136.1 C 221kPa 7:15  
VENTING 7:15  
CYCLE ABORTED 8:13  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 8:13  
-----

Digital Signature #  
91A2339650E4B8F5  
-----