
Midmark M11 - v1.0.0
Total Cycles: 3748

Name: RITTER 1
Sterilizer ID: V2178198
09/11/2023 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.1 F 0.0
2:00 189.4 F 2.5
4:00 199.6 F 7.0
6:00 215.3 F 12.0
8:00 223.5 F 6.3
10:00 234.6 F 11.1
12:00 232.8 F 8.0
14:00 243.2 F 11.9
16:00 245.4 F 12.8
18:00 257.4 F 19.2
20:00 268.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.3 F 28.7
2:00 271.6 F 28.4
3:00 272.3 F 28.7
4:00 271.4 F 28.2

Min 271.4 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 211.3 F 0.0
10:00 216.2 F 0.0
15:00 211.1 F 0.0
20:00 208.6 F 0.0
25:00 206.4 F 0.0
30:00 205.0 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 21:16
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:19
CYCLE COMPLETE

09/11/2023 11:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____